



Publication Year	2022
Acceptance in OA @INAF	2023-05-22T12:54:20Z
Title	Millimetric Sardinia radio Telescope Receiver based on Array of Lumped elements kids
Authors	D'Alessandro, G.; Barbarava, E.; Battistelli, E. S.; de Bernardis, P.; Cacciotti, F.; et al.
DOI	10.1051/epjconf/202225700012
Handle	http://hdl.handle.net/20.500.12386/34178
Journal	EPJ WEB OF CONFERENCES
Number	257

Millimetric Sardinia radio Telescope Receiver based on Array of Lumped elements kids

G. D'Alessandro^{1,2,*}, E. Barbarava¹, E.S. Battistelli^{1,2}, P. de Bernardis^{1,2}, F. Cacciotti¹, V. Capalbo¹, E. Carretti³, F. Columbro^{1,2}, A. Coppolecchia^{1,2}, A. Cruciani², M. De Petris^{1,2}, F. Govoni⁴, G. Isopi¹, L. Lamagna^{1,2}, P. Marongiu⁴, S. Masi^{1,2}, L. Mele^{1,2}, M. Molinari⁴, M. Murgia⁴, A. Navarrini⁴, A. Orlati³, A. Paiella^{1,2}, G. Pettinari⁵, F. Piacentini^{1,2}, T. Pisanu⁴, S. Poppi⁴, G. Presta^{1,2}, and F. Radiconi^{1,2}

¹Sapienza Università di Roma, Roma, Italy

²INFN sezione di Roma, 00185 Roma, Italy

³INAF - Istituto di Radioastronomia

⁴INAF - Osservatorio Astronomico di Cagliari

⁵CNR-IFN - Istituto di Fotonica e Nanotecnologie, Roma, Italy

Abstract. MISTRAL is a millimetric camera working in the W-band (78–103 GHz) which will take data from the Sardinia Radio Telescope, the Italian 64-m radio telescope located 50 km from Cagliari, at 600 m above the sea level, in Sardinia. It is being built as a facility instrument by the Sapienza University for INAF, that manages the radio telescope, under a PON contract. It will consist of a compact cryostat hosting the re-imaging optics, cooled at 4 K, and a 408-pixel array of photon-noise limited lumped element kinetic inductance detectors fabricated at CNR-IFN and cooled at a base temperature lower than 300 mK. MISTRAL will be able to investigate a long list of scientific targets spanning from extragalactic astrophysics to solar system science, with high angular resolution (~ 12 arcsec), including Sunyaev Zel'dovich effect measurements and the study of the Cosmic Web.

1 Introduction

Sardinia Radio telescope (SRT) is a multipurpose telescope aimed to measure the electromagnetic radiation from 300 MHz to 116 GHz. After the scientific commissioning, in 2015, 4 receivers have been installed and successfully operated at frequency lower than 26.5 GHz. New instruments are going to be installed in the next future, among which, MISTRAL will measure in the W-band.

High angular resolution W-band measurements open at a number of possibilities of scientific cases, including galaxies spectral energy distribution measurement and their continuum observation [1], AGN and radio galaxies [2], detection of circumstellar disks [3], dense core in giant molecular gas investigation [4], synchrotron non thermal jet and hot spot [5], intra-clusters filaments, cosmic web, and other S-Z related observation [6, 7]; and even more, by correlating with other experiment.

*e-mail: giuseppe.dalessandro@roma1.infn.it

Because of the large diameter telescope needed to perform high angular resolution measurements only few cameras around the world are devoted to this, such as MUSTANG2 [8] at the Green Bank Telescope.

MISTRAL aims to perform high angular resolution measurement with an array of lumped element kids detectors coupled, thanks to a re-imaging optic system, with SRT. MISTRAL is in an advanced phase of manufacturing and all the parts of the experiment are going to be assembled together at the beginning of 2022. MISTRAL will be installed in March 2022 and commissioning activities will start in spring 2022.

In section 2 we describe the Sardinia Radio Telescope enlightening its optical system and the typical observational conditions. Section 3 describes the instrument and its main parts.

2 Sardinia Radio Telescope

SRT [9] is a multipurpose instrument operated in either single dish or Very Long Baseline Interferometer mode. It is located 50km north of Cagliari in Sardinia (Italy), Lat. 39.4930N - Long. 9.2451E, 600 m above sea level. Construction was completed in August 2012. The technical commissioning phase to validate scientific performances was managed by National Institute for Astrophysics and concluded in 2014. The Early Science Program observations started in 2016, and regular observations in 2018.

The primary antenna (M1) is fully steerable, 64 m in diameter. It is composed of 1008 aluminum elements controlled by electromechanical actuators. An $f/0.33$ primary focus occurs near the sub-reflector (M2), 7.9 m in diameter, composed of 49 aluminum elements. Its position can be changed for focus adjustment. M1 and M2 are shaped to minimize spillover and the standing waves between the feed and the sub-reflector. The Gregorian focus, $f/2.34$ occurs around 20 meters below M2 in the Gregorian room, the violet box in Figure 1. MISTRAL will be placed in this room by using the Gregorian focus of SRT.

SRT has been designed to perform measurements from 305 MHz to 116 GHz [10]. Four receivers have been successfully installed: L-band (305 MHz - 410 MHz), P-band (1.3 GHz - 1.8 GHz), C-band (5.7 GHz - 7.7 GHz) and K-band (18 GHz - 26.5 GHz). In the next future, new receivers will increase the frequency coverage [10]: Q-band (33 GHz - 50 GHz), polarization-sensitive W-band (16 beams), K/Q/W band (3 beams) polarization-sensitive, and MISTRAL¹.

Table 1. SRT optical and mechanical main features

M1 diameter	64 m
M2 diameter	7.9 m
Primary focus	$f/0.33$
Gregorian focus	$f/2.34$
Pointing accuracy	2-13 arcsec
Range in elevation	5-90 deg
Range in azimuth	180 ± 270 deg

Among other activities, a metrological activity to measure with high precision deformations of the SRT due to gravitational and thermal effects is undergoing [10]; by means of the

¹The Enhancement of the Sardinia Radio Telescope (SRT) for the study of the Universe at high radio frequencies is financially supported by the National Operative Program (Programma Operativo Nazionale - PON) of the Italian Ministry of University and Research Research and Innovation 2014-2020, Notice D.D. 424 of 28-02-2018 for the granting of funding aimed at strengthening research infrastructures, in implementation of the Action II.1 – Project Proposal PIR01_00010




Figure 1. SRT schematic. The primary mirror (M1) consists of 1008 individual aluminum panels. Each panel has an area ranging from 2.4 to 5.3 m². It is built using aluminum sheets glued with epoxy resin. The sub-reflector (M2) consists of 49 individual aluminum panels. Near M2 occurs a primary f/0.33 focus. MISTRAL will be placed in the Gregorian room (violet box) where the f/2.34 Gregorian focus occurs.

active surface it will be possible to correct the deformations of the main mirror allowing the surface roughness needed for observations up to 116 GHz. Also, the metrological system will allow high precision pointing measuring and correcting the subreflector displacement from the optical axis together with the deformation of the whole structure.

2.1 Site condition

Estimation of sky opacity, based on recorded atmospheric data, forecasts by Navarrini et al. (<http://hdl.handle.net/20.500.12386/28787>) report $\tau < 0.15$ (50th percentile) at 93 GHz during the winter nights. The average PWV in the same conditions is mainly 8 mm. 50 years of radiosonde profiles taken at Cagliari airport (30 km far, at sea level) and scaled for SRT site shows PWV < 11 mm (50th percentile) and opacity < 0.2 (50th percentile) at 100 GHz [11]. What mainly concern W-band measurement a dynamic scheduling will be applied to optimize the observational time and weather conditions.

3 MISTRAL instrument

MISTRAL is a W-band camera for SRT. The focal plane is composed of 408 Lumped element kinetic inductance detectors (LEKID), fabricated at CNR-IFN, arranged on a 97 mm circular focal plane. The focal plane is optically coupled with SRT thanks to a re-imaging optics system composed of two silicon lenses coated with an anti-reflection system. The focal plane is cooled down at 270 mK with a cryostat provided by QMC instruments from Cardiff (UK).

3.1 Cryostat

The cryostat is provided by QMC instruments². It is composed of two radiation shields at 40 K and 4 K cooled down by a pulse tube cryocooler. Another shield, cooled at 1 K by a He⁴ fridge, surrounds the focal plane assembly. The detectors reach 270 mK thanks to the use of two He³ refrigerators. The external shell, made on Aluminum, is 700 mm diameter and 1680 mm high, see Figure 2. It is closed with a 190 mm UHMW-PE window [12]

The pulse tube is a Sumitomo RP-182B2S-F100H with 1.5 W at 4.2 K and 36 W at 48 K, and uses a remote valve in order to decrease the vibrations on the focal plane. The pulse tube head is mechanically disconnected to the shell thanks to a vibrations dumper. The pulse tube is water-cooled and we plan to use 100 m Helium lines[13].

The sub-kelvin stages are cooled down with a Chase Twin GL10 fridge: two He³ reaching 251 mK with loading of 20 μ W for the focal plane, a He³ reaching 332 mK with 30 μ W for the focal plane support and a He⁴ reaching 840 mK with 150 μ W for the focal plane shield.




Figure 2. MISTRAL cryostat receiver provided by QMC. A pulse tube refrigerator cooled down two radiation shields, at 40 K (orange box) and 4 K (green one). The fridges cooled down the 1 K shield (light blue box), the focal plane support and the focal plane (dark blue box). The focal plane is cooled down at ~ 270 mK

3.1.1 Magnetic shield

The experiment will move with the telescope during the observations. A magnetic shield surrounds the detectors, fridges, and relevant read-out parts to mitigate the Earth's magnetic field effects. The shield (1 mm thick) is made of Cryoperm 10 with $\mu_r > 70000$.

²<http://www.terahertz.co.uk/qmc-instruments-ltd>




Figure 3. MISTRAL optics chain. A UHMW window closes the vacuum shell with a clearance of 190 mm. The 40 K and 4 K shields are closed by a series of stack filters: for each shield, two thermal filters and a low pass edge filter. At 4 K the re-imaging optical system has been placed. The last low pass edge filter closes the 1 K shield. The last filter is a band-pass filter which defines the MISTRAL absorption band.

3.2 Optics

MISTRAL uses a re-imaging optical system, see Figure 3, to couple the Gregorian focus with the array focal plane. The re-imaging optics is composed of two Silicon lenses (L1 and L2) and a cold stop which is the aperture stop of the system. L1 is a convex-concave aspherical lens 290 mm in diameter. The cold stop is a 125 mm circular aperture coated with an absorber material (i.e. Eccosorb AN72). L2 is a convex-convex aspherical lens. The relay optics has been optimised to operate at 4 K assuming a Silicon refractive index of 3.4. Each lens surface will be covered with an antireflection coating based on one layer of PTFE; the efficiency in transmission is always $> 90\%$ over the MISTRAL frequency band. The field of view of the system is 4 arcmin by assuming an array focal plane 94.4 mm in diameter. The focal plane scale ratio is 2.54 arcsec/mm. The Strehl ratio on-axis is 0.97 with an FWHM of 12.2 arcsec; at the edge of the focal plane, the Strehl ratio decrease at 0.91, and the FWHM became 12.7 arcsec. The overall efficiency of MISTRAL optics chain is 90 % on-axis and decrease around 60 % near the focal plane edge.

3.3 Focal Plane

MISTRAL focal plane will be populated with 408 LEKIDs arranged on a triangular scheme with 4.2 mm side. Each absorber is a front-illuminated, 3rd order Hilbert, 3 mm \times 3 mm resonator, see Figure 4. The wafer is a 4 inches Silicon wafer 235 μ m thick, while the LEKIDs are made of superconducting Ti-Al bilayer (10 nm + 30 nm), as in [14, 15]. Due to the re-imaging optics two close-by pixels separation is 10.6 arcsec. The band-pass filter, placed in front of the wafer, closes the holder. It is centered at 90 GHz with ~ 25 GHz of bandwidth. The readout system is based on an FPGA, successfully used for OLIMPO [16].




Figure 4. *Left:* HFSS simulations have been performed to optimize the pixel absorption. *Right:* The 4 inches silicon wafer is anchored inside an aluminum holder thanks to beryllium-copper spring. The holder is closed with the band-pass filter and the continuity is guaranteed with SMA connectors.

3.3.1 Sensitivity estimation

To predict the performance on the sky, the measured noise equivalent power (NEP) has been compared with photon noise due to warm optics and atmospheric emission, Figure 4 in [17]. The latter, should be considered in its emission part and turbulence part which will act as correlated noise. A preliminary noise equivalent power about $3.5 \times 10^{-16} \text{ W} / \sqrt{\text{Hz}}$ has been measured at 270 mK by using a test-bed cryostat. By using experimental data, taken at SRT in the K-band, the atmospheric fluctuation noise, using modelling software such as *am*³, increases the atmospheric noise estimation a factor ≈ 10 with respect to the photon noise level even in the best 5% percentile resulting in a noise equivalent flux density of $\approx 3 \text{ mJy/beam}$.

4 Conclusions

SRT is a multipurpose observatory designed to measure a wide range of radio wavelengths: from 300 MHz to 116 GHz. At SRT, the sky opacity in winter is < 0.15 (50th percentile) at 93 GHz MISTRAL will be coupled with SRT with a re-imaging optical system. The minimum spatial resolution (FWHM) is 12.2 arcsec with a field of view of 4 arcmin. The 408 LEKIDs array has been optimized for the best 90 GHz absorption and for the background at SRT.

The scientific commissioning activities will start in summer 2022.

References

- [1] W.F. Wall, I. Puerari, R. Tilanus, F.P. Israel, J.E. Austermann, I. Aretxaga, G. Wilson, M. Yun, K.S. Scott, T.A. Perera et al., *Monthly Notices of the Royal Astronomical Society* **459**, 1440 (2016), 1603.07736
- [2] A. Humphrey, M. Zeballos, I. Aretxaga, D.H. Hughes, M.S. Yun, R. Cybulski, G.W. Wilson, J. Austermann, H. Ezawa, R. Kawabe et al., *Monthly Notices of the Royal Astronomical Society* **418**, 74 (2011), 1107.3120
- [3] M.S. Petersen, R.A. Gutermuth, E. Nagel, G.W. Wilson, J. Lane, *Monthly Notices of the Royal Astronomical Society* **488**, 1462 (2019)

³<https://www.cfa.harvard.edu/~spaine/am/>

- [4] A. Sokol, R. Gutermuth, R. Pokhrel, A. Gómez-Ruiz, G. Wilson, S. Offner, M. Heyer, A. Luna, F. Schloerb, D. Sánchez, *Monthly Notices of the Royal Astronomical Society* **483**, 407 (2019)
- [5] M.J. Hardcastle, L.W. Looney, *Monthly Notices of the Royal Astronomical Society* **388**, 176 (2008), [0804.3369](https://doi.org/10.1093/mnras/388.1.176)
- [6] T. Mroczkowski, D. Nagai, K. Basu, J. Chluba, J. Sayers, R. Adam, E. Churazov, A. Crites, L. Di Mascolo, D. Eckert et al., *Space Science Reviews* **215**, 17 (2019), [1811.02310](https://doi.org/10.1007/s11214-019-0631-0)
- [7] A. de Graaff, Y.C. Cai, C. Heymans, J.A. Peacock, *Astronomy & Astrophysics* **624**, A48 (2019), [1709.10378](https://doi.org/10.1051/0004-6361/201935738)
- [8] S.R. Dicker, P.A.R. Ade, J. Aguirre, J.A. Brevik, H.M. Cho, R. Datta, M.J. Devlin, B. Dober, D. Egan, J. Ford et al., *Journal of Low Temperature Physics* **176**, 808 (2014)
- [9] P. Bolli, A. Orlati, L. Stringhetti, A. Orfei, S. Righini, R. Ambrosini, M. Bartolini, C. Bortolotti, F. Buffa, M. Buttu et al., *Journal of Astronomical Instrumentation* **04**, 1550008 (2015), <https://doi.org/10.1142/S2251171715500087>
- [10] F. Govoni et al., 2021 XXXIVth General Assembly and Scientific Symposium of the International Union of Radio Science -, 1 (2021)
- [11] F.T. Nasir, F. Buffa, G.L. Deiana, *Experimental Astronomy* **29**, 207 (2011)
- [12] G. D'Alessandro, A. Paiella, A. Coppolecchia, M. Castellano, I. Colantoni, P. de Bernardis, L. Lamagna, S. Masi, *Infrared Physics & Technology* **90**, 59 (2018)
- [13] A. Coppolecchia, E. Battistelli, Masi et al., Pulse tube cooler with > 100m flexible lines optimized for operation of cryogenic detector arrays at large radiotelescopes (2021)
- [14] Catalano, A., Goupy, J., le Sueur, H., Benoit, A., Bourrion, O., Calvo, M., D'addabbo, A., Dumoulin, L., Levy-Bertrand, F., Macías-Pérez, J. et al., *Astronomy & Astrophysics* **580**, A15 (2015)
- [15] A. Paiella, A. Coppolecchia, M.G. Castellano, I. Colantoni, A. Cruciani, A. D'Addabbo, P. de Bernardis, S. Masi, G. Presta, *Journal of Low Temperature Physics* **184**, 97 (2016)
- [16] A. Paiella, E.S. Battistelli, M.G. Castellano, I. Colantoni, F. Columbro, A. Coppolecchia, G. D'Alessandro, P. de Bernardis, S. Gordon, L. Lamagna et al., *Journal of Physics: Conference Series* **1182**, 012005 (2019)
- [17] A. Coppolecchia, A. Paiella, L. Lamagna, G. Presta, E.S. Battistelli, P. de Bernardis, M.G. Castellano, F. Columbro, S. Masi, L. Mele et al., *Journal of Low Temperature Physics* **199**, 130 (2020)